

COMPARISON OF TWO METHODS FOR THE ANALYSIS OF PETIOLE NITRATE NITROGEN CONCENTRATION IN IRRIGATED COTTON

J.H. Smith, J.C. Silvertooth, and E.R. Norton
University of Arizona
Tucson, AZ

Abstract

A study was conducted in Arizona in 1997 with the objective of analyzing the accuracy of a recently developed portable nitrate meter (Cardy meter) to effectively measure nitrate-nitrogen ($\text{NO}_3\text{-N}$) in irrigated cotton (*Gossypium ssp.*). This task was accomplished by performing a correlation and linear regression analyses on $\text{NO}_3\text{-N}$ concentrations of cotton petiole sap, as measured by the Cardy meter, against the standard procedure using dried petiole $\text{NO}_3\text{-N}$ concentrations, as measured by the ion selective electrode (ISE). Results revealed that the $\text{NO}_3\text{-N}$ concentrations of petiole sap were highly correlated with dried petiole $\text{NO}_3\text{-N}$ (pearson correlation coefficient = 0.96, $P < 0.0001$). Therefore, a regression equation with an $r^2 = 0.92$ was derived: $Y = 9.96X - 1170.86$, where X and Y are $\text{NO}_3\text{-N}$ in petiole sap (ppm) and dried petiole (ppm), respectively. These results suggest that the sap analysis using the Cardy meter is a potentially valuable tool to monitor the in-season N status of irrigated cotton.

Introduction

Efficient nitrogen (N) management is essential to achieving optimum yields in cotton (*Gossypium spp.*) production. Due to the dynamic nature of cotton growth, decisions need to be made promptly and effectively regarding N rates and times of applications. For instance, N in excess of that needed for optimum lint yield can lead to excessive vegetative growth and delayed maturity (Tewolde et. al, 1995). Deficiencies experienced at the time of maximum crop demand can result in lint yield reductions. A technique available to managing in-season N inputs efficiently in terms of economic, agronomic, and environmental concerns can be accomplished by monitoring the in season $\text{NO}_3\text{-N}$ status of the plant using petiole analysis. Gardner and Tucker (1967) first established the foundation for the utilization of petioles as a tool when they identified a positive relationship between yield and petiole nitrate-N ($\text{NO}_3\text{-N}$) concentrations in irrigated cotton in the desert Southwest.

The traditional method of petiole analysis to determine plant N status involves numerous steps, which can be costly and time-consuming (Baker and Smith, 1969). A portable NO_3^- selective electrode (hereafter referred to as the Cardy meter) has recently been developed that can directly measure NO_3^-

present in fresh samples of expressed petiole sap. This Cardy meter offers immediate results of in-season crop N status. Therefore, adjustments in N fertilization can be made before the crop experiences N deficiencies or excessive N applications are made that may lead to enhanced vegetative growth, yield reductions and/or delayed maturity.

The accuracy of portable NO_3^- meters has been examined in the past by testing the correlation between $\text{NO}_3\text{-N}$ concentrations determined by sap analysis and $\text{NO}_3\text{-N}$ concentrations determined by dried petiole analysis. Hodges and Baker (1993) used cotton to compare the differences in the two methods. They found good correlations ($r = 0.88$) for early sampling periods, but poor correlations for late sampling periods. They found that the $\text{NO}_3\text{-N}$ concentrations measured in the petiole sap extracts increased relative to those measured in the dried petioles during later stages of growth, and concluded that this may be a result of the decrease in moisture content in mature cotton plants. Hartz et al. (1994) compared petiole sap $\text{NO}_3\text{-N}$ using a portable electrode with conventional dry tissue analysis for many vegetable crops and obtained r^2 values ranging from 0.65 to 0.89. Due to the slope of the regression line differing significantly among the crops, Hartz et al (1994) emphasized the importance of determining the relationship for each crop of interest.

Potential problems associated with the Cardy meter may result from poor calibration, environmental conditions, and maintenance. Initial experiments performed by Hodges and Baker, (1993) revealed that the meters were very sensitive to temperature and sunlight. Life of the meter's sensor pads will vary depending on environment and care. Initial measurements performed under laboratory conditions in this study revealed that the sensor pads may need replacement after approximately 260 samples.

In 1997, a study was conducted with the objective of analyzing the accuracy of the Cardy meter to effectively measure $\text{NO}_3\text{-N}$ in cotton under irrigated conditions in the desert Southwest.

Materials and Methods

Field experiments evaluating various rates of N applications were used as sampling sites in 1997. The experiments being conducted at these sites offered a wide gradient in petiole $\text{NO}_3\text{-N}$ concentrations. These experiments consisted of a N and PIX_{TM} (N-PIX) experiment conducted at the University of Arizona Marana Agricultural Center (MAR), and N management experiments at both the MAR and Maricopa Agricultural Center (MAC). Both Upland cotton (*G. hirsutum* L., var. DPL 33B) and American Pima (*G. barbadense* L., var Pima S-7) were sampled at MAC. At MAR, only Upland cotton, DPL 33B and STV 474, was sampled.

Routine petiole sampling from each experimental unit was performed on approximate 14 day intervals, from as early as the occurrence of the seventh node through cut-out. Petioles were sampled from the uppermost, fully expanded mature leaf, which generally fell in the vicinity of the fourth or fifth node below the terminal. Approximately 50 petioles were randomly collected from each plot, mixed thoroughly, and divided into two portions, dry and fresh. Fresh petioles were immediately stored in sealed plastic bags, and transported on ice in an insulated cooler from the field to the laboratory. The tissue of the fresh petioles was subsequently crushed with a rolling pin to extract sap, which was used for determining NO₃-N concentrations. Four to five drops of the extracted sap were placed directly on the sensor pad of a Cardy nitrate meter for analyses (Horiba, Ltd., Kyoto, Japan). Two Cardy meters were used simultaneously to help improve accuracy and precision of measurement. The Cardy meters were recalibrated using a two point calibration after every ten measurements or sooner if large differences between the Cardy meters were observed. Potassium nitrate (KNO₃) standards of 150 ppm NO₃⁻ and 2000 ppm NO₃⁻ were used to perform the calibrations. Nitrate measurements as determined by the Cardy meter were converted to NO₃-N.

Dry petioles were oven dried at 65 ° C for 24 hours and then ground to pass a 425 µm (40 mesh) screen for dry tissue analysis. Dried petiole samples were extracted with a 0.025 M Al₂(SO₄)₃ buffer solution containing 10 mg NO₃-N L⁻¹. The extraction process involved the addition of 40 ml of the extracting solution to 400 mg of sample and shaking the mixture for 15 minutes. Nitrate-N concentrations of the filtered extract were determined by use of an ion selective electrode (ISE) for NO₃-N. Regression and correlation analyses between the dry and fresh methods for NO₃-N determination were performed in accordance to procedures outlined by Gomez & Gomez (1984) and the SAS Institute (SAS, 1996).

Results and Discussion

Correlation and linear regression analyses were performed on NO₃-N concentrations of cotton petiole sap, as measured by the Cardy meter, against dried petiole NO₃-N, as measured by the ISE. The results revealed that NO₃-N concentrations of petiole sap were highly correlated with dried petiole NO₃-N throughout the season (pearson correlation coefficient = 0.96, P < 0.0001). The linear regression equation (Equation 1) was shown to be highly significant (P < 0.0001):

$$Y = 9.96X - 1170.86 \quad (n = 279, r^2 = 0.92) \quad (\text{Eq.1})$$

where Y refers to the NO₃-N in the dried petiole tissue (ppm), and X refers to the NO₃-N in the petiole sap (ppm). The standard errors of the slope and intercept are ± 0.0413 and ± 235, respectively. Both the slope and intercept are highly significant with P < 0.0001

In contrast to Hodges and Baker (1993), the NO₃-N concentrations in the petiole sap extracts measured in this study showed a consistent, positive linear relationship with the dried petiole measurements throughout the entire season (Fig.1). In this study, we conducted petiole sampling from the occurrence of the seventh node to cut-out, which provided samples with a range of 213.5 to 22, 485 ppm NO₃-N.

In conclusion, the results demonstrate that sap NO₃-N concentrations as determined by the Cardy meter were well correlated with dried petiole NO₃-N analyzed by the traditional laboratory method using dry, ground petioles. Therefore, the quick sap analysis using the Cardy meter is a potentially valuable tool to monitor the in-season N status of irrigated cotton in Arizona.

References

- Baker, A.S., and R. Smith. 1969. Extracting solution for potentiometric determination of nitrate in plant tissue. *J. Agric. Food Chem.* 17:1284-1287.
- Gardner, B.R., and T.C. Tucker. 1967. Nitrogen effects on cotton: II. Soil and petiole analysis. *Soil Sci. Soc. Am. Proc.* 31:785-791.
- Hartz, T.K., R.F. Smith, K.F. Schulbach, and M. LeStrange. 1994. On farm nitrogen tests improve fertilizer efficiency, protect groundwater. *Calif. Agri.* 48:29-32.
- Hodges, S.C., and S. Baker. 1993. Correlation of plant sap extracts of nitrate-N and K with dried petiole extracts. *Proc. Beltwide Cotton Conf. National Cotton Council, Memphis, TN.* p.1335-1337
- SAS Institute. 1996. SAS/STAT: Procedures. Release 6.03 ed. SAS Inst., Cary, NC.
- Sunderman, H.D., A.B. Onken, and L.R. Hossner. 1979. Nitrate concentration of cotton petioles as influenced by cultivar, row spacing, and N application rate. *Agron. J.* 71:p.731-737.
- Tewelde, H., C.J. Fernandez, D.C. Foss, and L.G. Unruh, Jr. 1995. Critical petiole nitrate-nitrogen for lint yield and enhanced maturity in pima cotton. *Agron. J.* 87:223-227.




Figure 1. Dry NO₃-N (ppm) vs. Sap NO₃-N (ppm)